

## RIGI- M1 Camera



# Uncooled real-time THz imager



- Beam profiling
- Non-destructive testing
- Imaging systems
- Quality control
- Security and surveillance
- Optical alignment

## Image using QCL THz source

Frequency: 4.6 THz
Power: 0.6 μ W
Integration time: 50 μs

• **Acquisition:** Single shot

	ltem	General Specifications
	Detector	Uncooled FPA THz micro-bolometer array
	Pitch	25 μm
	Array size	384 x 288 pixels
	Active area	9.6 x 7.2 mm
	Operating fr	equency 0.1 - 20 THz
	Pixel NEP	< 1.5 pW/√Hz at 4.6 THz
	Operating to	emperature -40 °C to 65 °C
	Software	Included and adaptable upon request

### Compatible with most terahertz sources

### 1. Broadband terahertz sources

- Broadband plasma based source
- Organic crystal high power spectrometer
- Lithium Niobate crystal high power spectrometer
- ZnTe crystal based linear spectrometer
- Photoconductive switch oscillator based source
- Commercial time domain spectrometer

### 2. Narrowband terahertz sources

- OCL based source
- Solid state sources
- Crystal based DFG source
- CO2-based DFG source

E-Mail: info@auniontech.com

Website: www.auniontech.com